

Bcl-X Ab-1 (Clone 2H12)

Mouse Monoclonal Antibody

Cat. #MS-715-P0, -P1, or -P (0.1ml, 0.5ml, or 1.0ml at 200µg/ml) (Purified Ab with BSA and Azide)

Cat. #MS-715-P1ABX or -PABX (0.1ml or 0.2ml at 1.0mg/ml) (Purified Ab without BSA and Azide)

Cat. #MS-715-B0, -B1, or -B (0.1ml, 0.5ml, or 1.0ml at 200µg/ml) (Biotin-labeled Ab with BSA and Azide)

Cat. #MS-715-R7 (7.0ml) (Ready-to-Use for Immunohistochemical Staining)

Cat. #MS-715-PCS (5 Slides) (Positive Control for Histology)

Cat. #PP-715 (1.0ml at 200µg/ml) (Blocking peptide with BSA and Azide)

Description: Bcl-X has two isoforms, Bcl-X_L (long) a 241 amino acid protein and Bcl-X_S (short) a 178 amino acid protein lacking a 63 amino acid domain that is well conserved among members of the Bcl-2 family.

Comments: Ab-1 reacts with both Bcl-X_S and Bcl-X_L proteins.

Mol. Wt. of Antigen: 27kDa

Epitope: aa 3-14

Species Reactivity: Human¹, Mouse¹, Rat¹, and Pig. Others-not tested.

Clone Designation: 2H12

Ig Isotype: IgG_{2a}

Immunogen: A synthetic peptide, aa 3-14 (Cys-QSNRELVVDFLS) of human Bcl-X protein.¹ This amino acid sequence is shared by human and murine bcl-X protein.¹

Applications and Suggested Dilutions:

- Flow Cytometry
- Immunofluorescence³
- Western Blotting (Not optimum)
- Immunohistology (Formalin/paraffin)
(Use Ab 2-4µg/ml for 30 min at RT)
- * [Staining of formalin-fixed tissues REQUIRES boiling tissue sections in 10mM citrate buffer, pH 6.0, (**NEOMARKERS'** Cat. #AP-9003), for 10-20 min followed by cooling at RT for 20 min.]

The optimal dilution for a specific application should be determined by the investigator.

Positive Control: Hodgkin's lymphoma.

Cellular Localization: Cytoplasmic and cell/nuclear membrane.

Supplied As:

200µg/ml antibody purified from the ascites fluid by Protein A chromatography. Prepared in 10mM PBS, pH 7.4, with 0.2% BSA and 0.09% sodium azide. Also available without BSA and azide at 1mg/ml,

or

Prediluted antibody which is ready-to-use for staining of formalin-fixed, paraffin-embedded tissues.

Storage and Stability:

Ab with sodium azide is stable for 24 months when stored at 2-8°C. Antibody WITHOUT sodium azide is stable for 36 months when stored at below 0°C.

Key References:

1. Hsu YT, et. al. J Bio Chem, 1997, 272:13829-34.
2. Hsu YT, et. al. PNAS USA, 1997, 94(8):3668-72.
3. Wolter KG, et. al. J Cell Bio, 1997, 139(5):1281-92.

Limitations and Warranty:

Our products are intended FOR RESEARCH USE ONLY and are not approved for clinical diagnosis, drug use or therapeutic procedures. No products are to be construed as a recommendation for use in violation of any patents. We make no representations, warranties or assurances as to the accuracy or completeness of information provided on our data sheets and website. Our warranty is limited to the actual price paid for the product. NeoMarkers is not liable for any property damage, personal injury, time or effort or economic loss caused by our products.

Material Safety Data:

This product is not licensed or approved for administration to humans or to animals other than the experimental animals. Standard Laboratory Practices should be followed when handling this material. The chemical, physical, and toxicological properties of



Bcl-X Ab-1 (Clone 2H12)**Mouse Monoclonal Antibody**

Cat. #MS-715-P0, -P1, or -P (0.1ml, 0.5ml, or 1.0ml at 200µg/ml) (Purified Ab with BSA and Azide)

Cat. #MS-715-P1ABX or -PABX (0.1ml or 0.2ml at 1.0mg/ml) (Purified Ab without BSA and Azide)

Cat. #MS-715-B0, -B1, or -B (0.1ml, 0.5ml, or 1.0ml at 200µg/ml) (Biotin-labeled Ab with BSA and Azide)

Cat. #MS-715-R7 (7.0ml) (Ready-to-Use for Immunohistochemical Staining)

Cat. #MS-715-PCS (5 Slides) (Positive Control for Histology)

Cat. #PP-715 (1.0ml at 200µg/ml) (Blocking peptide with BSA and Azide)

this material have not been thoroughly investigated. Appropriate measures should be taken to avoid skin and eye contact, inhalation, and ingestion. The material contains 0.09% sodium azide as a preservative. Although the quantity of azide is very small, appropriate care should be taken when handling this material as indicated above. The National Institute of Occupational Safety and Health has issued a bulletin citing the potential explosion hazard due to the reaction of sodium azide with copper, lead, brass, or solder in the plumbing systems. Sodium azide forms hydrazoic acid in acidic conditions and should be discarded in a large volume of running water to avoid deposits forming in metal drainage pipes.

For Research Use Only

Thermo Fisher Scientific
Anatomical Pathology
47777 Warm Springs Blvd.
Fremont, CA 94539, USA
Tel: 1-510-991-2800
Fax: 1-510-991-2826
<http://www.thermo.com/labvision>



Manufactured by:
NeoMarkers
For
Lab Vision Corporation



Thermo Fisher Scientific
Anatomical Pathology
93-96 Chadwick Road, Astmoor
Runcorn, Cheshire WA7 1PR, UK
Tel: 44-1928-562600
Fax: 44-1928-562627
Labvision.uk@thermofisher.com